

3M Poland Manufacturing Sp. z o.o  
ul. Kwidzyńska 6  
51-416 Wrocław  
Tel.: +48 71 325 25 52  
Fax: +48 71 325 31 18  
www.3m.pl



Document version: 1

## ***Declaration of Conformity***

We

**3M Poland Manufacturing Sp. z o.o.  
ul. Kwidzyńska 6  
51-416 Wrocław  
Poland**

hereby declare under our sole responsibility that the CE marked products,  
to which this declaration relates,

**Orthoptic eye patches:**

### **OPTICLUDE**

1537/20, 1537/50, 1537/100, 1537M-50, 1537M-100, 1537T  
2537PE, 2537PF, 2537MPE  
2537M D PB-50, 2537M D PG-50, 2537M D PB-100, 2537M D PG-100,  
1538M-50, 1538M-100, 2538PE, 2538MPE, 2538 MPB-50  
2538M D PB-50, 2538M D PG-50, 2538M D PB-100, 2538M D PG-100, 2538M P-50 Disney Edition NL,  
1539/20, 1539/50, 1539/100, 1539M-50, 1539M-100, 1539T,  
2539PE, 2539PF, 2539MPE  
2539M D PB-50, 2539M D PG-50, 2539M D PB-100, 2539M D PG-100, 2539M P-50 Disney Edition NL,  
1637M D PG-100, 1637M D PB-100, 1643B, 1643E/20, 1643G, 1643V, 1643VF, 1643MP  
1638M D PG-100, 1638M D PB-100, 1638M D PG-50, 1638M D PB-50,  
1646 MV-50, 1646B, 1646E/20, 1646F/20, 1646G, 1646V, 1646VF, 1646MP  
1639M D PG-100, 1639M D PB-100  
1648B, 1648E/20, 1648G, 1648V, 1648VF, 1648MP


### **Nexcare OPTICLUDE**

NE-OPT6250, NE-OPT8257, 1537, 1539

are classified, according to rule 4 of Annex IX of the  
Medical Device Directive 93/42/EEC, as **Class I** non-sterile devices  
and are in accordance with

***Annex VII and all other applicable provisions of the Directive 93/42/EEC***  
on the approximation of the laws of the European Union Member States  
concerning medical devices.

Signature: \_\_\_\_\_

  
Dominika Kawala  
Manufacturing Quality Manager

Data: \_\_\_\_\_

18.06.2013

relating to REG-STED-05-131830